

ई आई भवन, 5वां तल, प्लॉट सं. 85, सेक्टर 11, खारघर, नवी मुंबई- 410210
El Bhawan, 5th Floor, Plot No 85, Sector-11, Kharghar, Navi Mumbai-410210

Report. No.: EIL/MUM/4866/2025-26/AKC/015

To Whomsoever It May Concern

Sub: Calibration Witness Report of UTM (Digital) at M/s. Global Metascan Analysis LLP.

This is to certify that at the request of M/s Global Metascan Analysis LLP. vide their email dated 28.01.2026, the undersigned paid a visit to their Laboratory at Plot No. PAP-R-534, 535, R-Block, T.B. Road, Near Apar Industries Gate No. 3, TTC Industrial Area, MIDC, Rabale, Navi Mumbai Maharashtra to witness calibration of Universal Testing Machine (Digital) referred hereunder:

Model	:	UTE-40T (400 KN) SERVO CONTROL
Sr. No.	:	173-1225
Capacity	:	400 KN
Calibration Range	:	5-400 KN
Make	:	MCS

The Machine was calibrated in compression mode and verified with calibrated load cell Sr. No. JX2103238, capacity 100 KN and Sr. No. 1797698 capacity 1000 KN by M/s Size Master Calibration Services, Pune. The results of the machine were found to be within the permissible limits of $\pm 1\%$ accuracy/ error as per **Class I** machine range of Indian Standard **IS 1828 (Part-I):2022 & ISO 7500-I:2018** (Refer Certificate No. 2026 / 78547 / 03 dated 06.02.2026 issued by M/s. Size Master Calibration Services) and the accuracy error & repeatability error found within limit $\pm 1\%$ for the verified range as per clause No. 16.1 of **ASTM E-4:2024** (Refer Certificate No. 2026 / 78547 / 04 dated 06.02.2026 issued by M/s. Size Master Calibration Services).

Date of Calibration : 5th February 2026

Note: The above testing machine would become due for next calibration on **4th February 2027** or earlier if the machine is reinstalled at any other location or overhauled or repaired or there is any reason to doubt its accuracy/precision prior to this date(s).

Date of Issue: 16th February 2026

Inspection Authority

पंजीकृत कार्यालय : ईआईएल भवन, 1, भीकाजी कामा प्लेस, नई दिल्ली-110 066 | Regd. Office : EIL Bhawan, 1, Bhikaji Cama Place, New Delhi-110 066 | फोन / Phone : 011-26762121 (EPABX)

Celebrating
60
YEARS
of Nation Building
(Since 1965)

YEAR OF
TRANSFORMATION

स्थानीय कार्यालय - खारघर/Kharghar फोन/ Phone : 022 27740051, 27524801

उत्कृष्टता का आधार - हमारे कर्म **Delivering Excellence through People**

CIN : L74899DL1965GOI004352

Website : www.engineersindia.com